

August 14, 2018

Mott Community College
1401 East Court St.
Flint, MI 48503

Subject: Mott Community College
20180807 MCC Sites

Dear Ms. Hunt :

Thank you for making Brighton Analytical, L.L.C. your laboratory of choice. Attached are the results for the samples submitted on 08/07/2018 for the above mentioned project. NELAP/TNI Accredited Analysis and MDEQ Drinking Water Certified Analysis will be identified in their respective reporting formats. Hard copies can be supplied at your request for a fee of \$20.00 per copy.

The invoice for this project will be emailed separately. If you have any questions concerning the data or invoice, please don't hesitate to contact our office. We welcome your comments and suggestions to improve our quality systems. Please reference Brighton Analytical, L.L.C. Project ID 52092 when calling or emailing. We thank you for this opportunity to partner with you on this project and hope to work with you again in the future.

Sincerely,
Brighton Analytical, L.L.C.



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 07:54
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00253**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LAPR_FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:01	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 07:55
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00254**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LAPR_S 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:06	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 07:57
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00255**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LAPR_S 1-3**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	4	ug/L	1	15	EPA 200.8 rev5.4	13:10	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 07:58
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00256**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LAPR_HS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:15	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 07:59
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00257**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LAPR_S 1-4**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:19	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 08:01
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00258**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LAPR_S 1-5**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------


Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:24	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 08:02
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00259**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LAPR_HS 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:29	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 09:01
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00260**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **NTC_S 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:33	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 09:03
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00261**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **NTC_S 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:52	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 09:05
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00262**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **NTC_DF 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	13:56	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 09:05
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00263**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **NTC_DF 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	14:14	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 10:07
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00264**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **GLBC_FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------


Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	14:18	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 10:28
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00265**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **HTC_FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	14:23	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 10:23
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00266**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **HTCFP_DF 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	2	ug/L	1	15	EPA 200.8 rev5.4	14:28	08/13/2018
-----------------------------	---	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 10:24
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00267**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **HTCFP_DF 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	2	ug/L	1	15	EPA 200.8 rev5.4	14:32	08/13/2018
-----------------------------	---	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:04
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00268**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_111 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	14:37	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 10:58
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00269**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_110 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------


Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	14:41	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:05
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00270**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_109 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:00	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:11
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00271**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_108 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:04	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:12
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00272**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_106 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------


Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:09	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:14
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00273**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_107 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:40	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:15
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00274**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_105 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:45	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:16
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00275**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_103 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:49	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:16
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00276**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_104 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:54	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:17
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00277**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_102 FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	15:58	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:19
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00278**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_101 S 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------


Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	16:17	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:09
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00279**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **ECD_HS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	16:21	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:39
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00280**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **DN_FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------


Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	16:26	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 11:38
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00281**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **DN_HS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1630	15	EPA 200.8 rev5.4	16:26	08/13/2018
-----------------------------	--------------	------	------	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:24
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00282**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_FS 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	16:35	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:28
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00283**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_DF 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	16:53	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:36
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00284**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_DF 1-3**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	1	ug/L	1	15	EPA 200.8 rev5.4	16:57	08/13/2018
-----------------------------	---	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:39
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00285**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_S 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	17:02	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:44
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00286**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_S 1-3**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------


Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	17:06	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:47
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00287**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_FS 1-4**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	17:25	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:48
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00288**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_DF 1-4**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	17:29	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:41
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00289**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_S 1-5**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	17:34	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:42
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00290**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_S 1-6**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	2	ug/L	1	15	EPA 200.8 rev5.4	17:38	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:54
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00291**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB_DF 1-5**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	1	ug/L	1	15	EPA 200.8 rev5.4	17:43	08/13/2018
-----------------------------	---	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:17
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00292**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLC_DF 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	17:47	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 13:39
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00293**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LIV_S 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	18:19	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 13:40
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00294**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LIV_S 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	18:23	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 13:40
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00295**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LIV_S 1-3**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	18:41	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 13:41
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00296**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LIV_DF 1-1**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	18:46	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 13:42
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00297**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LIV_DF 1-2**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	18:51	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 13:52
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00298**


Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LIV_DF 1-3**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	18:55	08/13/2018

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
 Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 13:53
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00299**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **LIV_DF 1-4**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	19:00	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, Michigan 48114
 Phone: (810)229-7575 (810)229-8650
 e-mail: bai-brighton@sbcglobal.net
 MDNRE Certified #9404
 NELAC Accredited #176507

Sample Date/Time: 08/07/2018 12:45
 Submit Date/Time: 08/07/2018 14:30
 Report Date: 08/14/2018

Mott Community College
 1401 East Court St.
 Flint, MI 48503

BA Project # **52092**
 BA Sample ID **CI00300**

Project Name: **Mott Community College**
 Project Number: **20180807 MCC Sites**
 Sample ID: **SLB S 1-7**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
--------------	--------	-------	----	-----	------------------	---------------	---------------

Drinking Water Metal Analysis

Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	19:04	08/13/2018
-----------------------------	--------------	------	---	----	------------------	-------	------------

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 

Date 8/14/2018

BA PROJECT #: 52022		Analysis Requested/Method		1 PAGE OF 5	
ABBREVIATIONS FOR MATRIX S = Solid L = Liquid DW = Drinking H ₂ O WW = Wastewater O = Oil P = Wipe A = Air (Tedar Bag) F = Filter T = Tube M = Methanol		REPORT RESULTS TO: Mott Comm College			
COMPANY NAME: Brighton Analytical LLC 2105 Pless Drive Brighton, MI 48114 Phone: 810-229-7575 Fax: 810-229-8650		Attn:			
PROJECT NAME: Mott Community College		PHONE/FAX:			
PROJECT NUMBER: 20180807 MCC Sites		Sample received within holding time? yes <input checked="" type="checkbox"/> no <input type="checkbox"/>			
P. O. NUMBER:		For TCLP ONLY — Federal Limits J Other 0			
Sample Collected by: Tamara Hunt		Samples intact: yes <input checked="" type="checkbox"/> no <input type="checkbox"/> (if no, see below)			
REQUESTED TURNAROUND: (circle one) Rush: 1-3 business days (verify with lab & specify date needed) Expedited: 5 business days Standard: 10 business days		Note samples if not intact:			
Brighton ID #	Sample Description	Time	Date	Temp	
1) 253	LAPR_FS 1-1	0754	7-18-18		
2) 54	LAPR_S 1-2	0755			
3) 55	LAPR_S 1-3	0757			
4) 56	LAPR_HS 1-1	0758			
5) 57	LAPR_S 1-4	0759			
6) 58	LAPR_S 1-5	0801			
7) 59	LAPR_HS 1-2	0802			
8) 60	NTC_S 1-1	0806			
9) 61	NTC_S 1-2	0903			
10) 62	NTC_DF 1-1	0905			
Headspace/bubbles in VOAS? yes 0 no 1/n/a 0		Sample containers and COC match? yes <input checked="" type="checkbox"/> no <input type="checkbox"/>			
Comments:					
Temperature of Samples °C:					

Special Instructions: Please fill out the Chain of Custody completely and review. Incorrect or incomplete information will result in a "hold" on all analyses.

Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:	Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:
1			8-7-18	1430	3				



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, MI 48114 Phone: 810-229-7575
 Fax: 810-229-8650

COMPANY NAME: Mott Community College
PROJECT NAME: 20180807 MCC Sites
PROJECT NUMBER:

P. O. NUMBER:

Sample Collected by: Tamara Hunt
REQUESTED TURNAROUND: (circle one)
 Rush: 1-3 business days (verify with lab & specify date needed) Expedited: 5 business days
 Standard: 10 business days

IF RUSH, approved by:

Sampling Time Date

Brighton ID #	Sample Description	Time	Date
263	NTC_DF 1-2	0905	7/18
264	GLBC_FS 1-1	1007	7/18
265	HTC_FS 1-1	1028	7/18
266	HTCFP_DF 1-1	1033	7/18
267	HTCFP_DF 1-2	1024	7/18
268	ECD_111 FS 1-1	1104	7/18
269	ECD_110 FS 1-1	1058	7/18
270	ECD_109 FS 1-1	1105	7/18
271	ECD_108 FS 1-1	1111	7/18
272	ECD_106 FS 1-1	1112	7/18

Special Instructions:

BA PROJECT #: 5202

ABBREVIATIONS FOR MATRIX

S = Solid
 L = Liquid
 DW = Drinking H₂O
 WW = Wastewater
 O = Oil
 P = Wipe
 A = Air (Tedar Bag)
 F = Filter
 T = Tube
 M = Methanol

Container Type & Quantity

VOALS (PRES) Y/N	HOPE (PRESERVED)	HOPE (NO)	HOPE (NO)	HOPE (NO)	HOPE (NO)	GLASS, NO PRESERVATIVE	GLASS (H.S.O)	MEOH Preserved	(Field or Lab Preserved)
	X								
	X								
	X								
	X								
	X								
	X								
	X								
	X								
	X								
	X								

Sample Matrix

Total Lead

Temp

Temperature of Samples °C:

Sample received within holding time? yes no

For TCLP ONLY — Federal Limits J Other 0

Samples intact? yes no (if no, see below)

Note samples if not intact:

Headspace/bubbles in VOCs? yes 0 no 0 n/a 0

Sample containers and COC match? yes no

Comments:

Analysis Requested/Method

REPORT RESULTS TO:

2 PAGE OF

Attn:

PHONE FAX:

Please fill out the Chain of Custody completely and review. Incorrect or incomplete information will result in a "hold" on all analyses.

Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:	Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:
1			8.7.11	1430	3				

BA
Brighton Analytical LLC
 2105 Pless Drive
 Brighton, MI 48114 Phone: 810-229-7575
 Fax: 810-229-8650

COMPANY NAME: Mott Community College
PROJECT NAME: 20180807 MCC Sites
PROJECT NUMBER:

P. O. NUMBER:

Sample Collected by: Tamara Hunt
REQUESTED TURNAROUND: (circle one)
 Rush: 1-3 business days (verify with lab & specify date needed) Expedited: 5 business days
 Standard: 10 business days

Brighton ID #	Sample Description	Sampling		Container Type & Quantity								Sample Matrix	Analysis Requested/Method	
		Time	Date	VOAS (PRES Y/N)	HDPE (PRESERVED)	HDPE (INFO)	HDPE (NAOH)	ALVIBET	GLASS, NO PRESERVATIVE	MEOH Preserved (Field or Lab Preserved)				
1) 273	ECD_107 FS 1-1	1114	7/16/18		X									
2) 74	ECD_105 FS 1-1	1115		X										
3) 75	ECD_103 FS 1-1	1116		X										
4) 76	ECD_104 FS 1-1	1116		X										
5) 77	ECD_102 FS 1-1	1117		X										
6) 78	ECD_101 FS 1-1	1119		X										
7) 79	ECD_HS 1-1	1109		X										53.8
8) 80	DN_FS 1-1	1130		X										
9) 81	DN_HS 1-1	1138		X										
10) 82	SLB_FS 1-1	1204		X										

Special Instructions: Temperature of Samples °C:

Please fill out the Chain of Custody completely and review. Incorrect or incomplete information will result in a "hold" on all analyses.

Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:	Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:
1			8-7-18	1430	3				

REPORT RESULTS TO:

Attn:
PHONE: FAX:

Sample received within holding time? yes no

For TCLP ONLY — Federal Limits Other 0

Samples intact: yes no (if no, see below)

Note samples if not intact:

Headpace/bubbles in VOAS? yes 0 no 0 n/a 0

Sample containers and COC match? yes no

Comments:

Temperature of Samples °C:

Kaye
Adjust total
acc

BA

Brighton Analytical LLC
2105 Pless Drive
Brighton, MI 48114 Phone: 810-229-7575
Fax: 810-229-8650

COMPANY NAME: **Mott Community College**
PROJECT NAME: **20180807_MCC Sites**
PROJECT NUMBER:

BA PROJECT #: **5202**

ABBREVIATIONS FOR MATRIX:
S = Solid
L = Liquid
DW = Drinking H₂O
WW = Wastewater
O = Oil
P = Wipe
A = Air (Tractor Bag)
F = Filter
T = Tube
M = Methanol

4 PAGE OF
REPORT RESULTS TO:

Attn:
PHONE: FAX:

Sample received within holding time? yes no

For TCLP ONLY — Federal Limits Other 0

Samples intact: yes no (if no, see below)

Note samples if not intact:

Headspace/bubbles in VOAS? yes 0 no 0 n/a 0

Sample containers and COC match? yes no

Comments:



Temperature of Samples °C:

Brighton ID #	Sample Description	Time	Date	VOAS (PRES) Y/N			Container Type & Quantity						Temp			
				HEADSPACE	SEALS	IDENTIFIED	TYPE	QTY	TYPE	QTY	TYPE	QTY		TYPE	QTY	
1)	SLB_DF 1-1	1228	7/18													
2)	SLB_DF 1-3	1236	7/18													
3)	SLB_S 1-2	1239	7/18													
4)	SLB_S 1-3	1244	7/18													
5)	SLB_FS 1-4	1247	7/18													
6)	SLB_DF 1-4	1248	7/18													
7)	SLB_S 1-5	1241	7/18													
8)	SLB_S 1-6	1242	7/18													
9)	SLB_DF 1-5	1244	7/18													
10)	SLC_DF 1-1	1247	7/18													

Special Instructions:

Temperature of Samples °C:

Please fill out the Chain of Custody completely and review. Incorrect or incomplete information will result in a "hold" on all analyses.

Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:	Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:
1			8-7-18	1430	3				



Brighton Analytical LLC
 2105 Pless Drive
 Brighton, MI 48114 Phone: 810-229-7575
 Fax: 810-229-8650

COMPANY NAME: Mott Community College
PROJECT NAME: 20180807_MCC Sites
PROJECT NUMBER:

P. O. NUMBER:

Sample Collected by: Tamara Hunt
REQUESTED TURNAROUND: (circle one)
 Rush: 1-3 business days (verify with lab & specify date needed) Expedited: 5 business days
 Standard: 10 business days

IF RUSH, approved by:
Sampling Time Date

Brighton ID #	Sample Description	Sampling Time	Date	VOAS PRES Y/N	THREE PRESERVED	TOP TINO.	THREE T.S.O.	THREE NAOH	ALIBETH	GLASS, NO PRESERVATIVE	MEOH Preserved (Field or Lab Preserved)
1) 293	LIV_S 1-1	1339	8/6/08		X						
2) 294	LIV_S 1-2	1340			X						
3) 295	LIV_S 1-3	1340			X						
4) 296	LIV_DF 1-1	1341			X						
5) 297	LIV_DF 1-2	1342			X						
6) 298	LIV_DF 1-3	1352			X						
7) 299	LIV_DF 1-4	1353			X						
8) 300	SIB 51-7	1845			N						
9)											
10)											

Analysis Requested/Method
 Sample Matrix
 Total Lead
 Temp
 47.5
 77.6
 Temperature of Samples °C:

BA PROJECT #: 52052
REPORT RESULTS TO:

ABBREVIATIONS FOR MATRIX
 S = Solid
 L = Liquid
 DW = Drinking H₂O
 WW = Wastewater
 O = Oil
 P = Wipe
 A = Air (radial Bag)
 F = Filter
 T = Tube
 M = Methanol

Sample received within holding time? yes no

For TCLP ONLY — Federal Limits Other 0

Samples intact: yes no (if no, see below)

Note samples if not intact:

Headspace/bubbles in VOAS? yes 0 no 0 n/a 0

Sample containers and COC match? yes no

Comments:

Please fill out the Chain of Custody completely and review. Incorrect or incomplete information will result in a "hold" on all analyses.

Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:	Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:
1			8-7-08	1430	3				



BRIGHTON ANALYTICAL, LLC

QUALITY ASSURANCE/QUALITY
CONTROL

ICP-MS

METHOD 200.8/6020

REPRESENTATIVE BATCH PRECISION AND ACCURACY QUALITY CONTROL SUMMARY

Analysis Date: 08/13/2018

Standard ID: 071318 H20

Batch: 8/9/2018 B1

Matrix Spike Lab ID: CI00262

Matrix: Total

Analyst: LT

Metals	Matrix Spike - Precision *			Matrix Spike - Accuracy**				Miscellaneous***		
	Matrix Spike (ug/kg)	Matrix Spike Dup (ug/kg)	RPD (%)	Spk Conc (ug/kg)	MS Recovery (%)	MSD Recovery (%)	Sample Conc (ug/kg)	Method Blk (ug/kg)	LCS-Method STD (%)	Ind. Std. (%)
Lead	1080	1082	0.2	1000	108.0	108.2	0	<1	103.9	101.0

* Matrix spike precision range +/- 20% RPD

** Matrix spike accuracy range +/- 20% recovery

*** LCS accuracy range +/- 15% recovery / Ind std accuracy range +/- 10% recovery

Comments: _____

ICP-MS

METHOD 200.8/6020

REPRESENTATIVE BATCH PRECISION AND ACCURACY QUALITY CONTROL SUMMARY

Analysis Date: 08/13/2018

Standard ID: 071318 H20

Batch: 8/9/2018 B2

Matrix Spike Lab ID: CI00282

Matrix: Total

Analyst: LT

Metals	Matrix Spike - Precision *			Matrix Spike - Accuracy**				Miscellaneous***		
	Matrix Spike (ug/kg)	Matrix Spike Dup (ug/kg)	RPD (%)	Spk Conc (ug/kg)	MS Recovery (%)	MSD Recovery (%)	Sample Conc (ug/kg)	Method Blk (ug/kg)	LCS-Method STD (%)	Ind. Std. (%)
Lead	1057	1068	1.0	1000	105.7	106.8	0	<1	102.8	101.0

* Matrix spike precision range +/- 20% RPD

** Matrix spike accuracy range +/- 20% recovery

*** LCS accuracy range +/- 15% recovery / Ind std accuracy range +/- 10% recovery

Comments: _____

ICP-MS

METHOD 200.8/6020

REPRESENTATIVE BATCH PRECISION AND ACCURACY QUALITY CONTROL SUMMARY

Analysis Date: 08/13/2018

Standard ID: 071318 H20

Batch: 8/9/2018 B3

Matrix Spike Lab ID: CI00300

Matrix: Total

Analyst: LT

Metals	Matrix Spike - Precision *			Matrix Spike - Accuracy**				Miscellaneous***		
	Matrix Spike (ug/kg)	Matrix Spike Dup (ug/kg)	RPD (%)	Spk Conc (ug/kg)	MS Recovery (%)	MSD Recovery (%)	Sample Conc (ug/kg)	Method Blk (ug/kg)	LCS-Method STD (%)	Ind. Std. (%)
Lead	1054	1044	1.0	1000	105.4	104.4	0	<1	109.2	101.0

* Matrix spike precision range +/- 20% RPD

** Matrix spike accuracy range +/- 20% recovery

*** LCS accuracy range +/- 15% recovery / Ind std accuracy range +/- 10% recovery

Comments: _____